

Location :	Main
MC	NW
NE	OPS

Refrigerator/Freezer Temperature Log

Circle one

Month/Year:			Freezer	Vaccine	Allergy	Infusion
Day of the month	AM temp in Celsius	Alarm on? Yes or no	Initials	PM temp in Celsius	Alarm on? Yes or no	Initials
1						
2						
3						
4						
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						
31						

Staff Initials/Full Signature & Credentials:

_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____

Temp Ranges- Refrigerator: 2°C-8°C, Freezer: -45°C- -20°C

- Temperature monitoring and recording should occur 2 times each work day with temp logged and visual inspection of the unit.
- Rotate stock so that vaccines closest to their expiration date will be used first.
- Remove expired vaccines and diluents dispose of properly, we do not want them administered to patients.
- **RESPOND IMMEDIATELY** to Sharla Smith (931-220-3317), Tiffany Hampton (931-241-7378) or Michelle Smith (931-220-4346) via telephone when potential temperature excursions have occurred or if there is a possibility of equipment malfunction.